



UWL REPOSITORY

repository.uwl.ac.uk

Automated assessment of transthoracic echocardiogram image quality using deep neural networks

Labs, Robert B., Vrettos, Apostolos, Loo, Jonathan and Zolgharni, Massoud ORCID logo ORCID:
<https://orcid.org/0000-0003-0904-2904> (2022) Automated assessment of transthoracic echocardiogram image quality using deep neural networks. *Intelligent Medicine*. ISSN 2667-1026

<http://dx.doi.org/10.1016/j.imed.2022.08.001>

This is the Published Version of the final output.

UWL repository link: <https://repository.uwl.ac.uk/id/eprint/9396/>

Alternative formats: If you require this document in an alternative format, please contact: open.research@uwl.ac.uk

Copyright: Creative Commons: Attribution 4.0

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

Take down policy: If you believe that this document breaches copyright, please contact us at open.research@uwl.ac.uk providing details, and we will remove access to the work immediately and investigate your claim.

Rights Retention Statement:



Research Article

Automated assessment of transthoracic echocardiogram image quality using deep neural networks

Robert B. Labs^{1,*}, Apostolos Vrettos², Jonathan Loo¹, Massoud Zolgharni¹

¹ School of Computing and Engineering, University of West London, London, United Kingdom

² Imperial College, Healthcare, NHS Trust, United Kingdom



ARTICLE INFO

Keywords:

Image quality
Echocardiography
Objective assessment
Deep learning
Ultrasound

ABSTRACT

Background Standard views in two-dimensional echocardiography are well established but the qualities of acquired images are highly dependent on operator skills and are assessed subjectively. This study was aimed at providing an objective assessment pipeline for echocardiogram image quality by defining a new set of domain-specific quality indicators. Consequently, image quality assessment can thus be automated to enhance clinical measurements, interpretation, and real-time optimization.

Methods We developed deep neural networks for the automated assessment of echocardiographic frames that were randomly sampled from 11,262 adult patients. The private echocardiography dataset consists of 33,784 frames, previously acquired between 2010 and 2020. Unlike non-medical images where full-reference metrics can be applied for image quality, echocardiogram's data are highly heterogeneous and requires blind-reference (IQA) metrics. Therefore, deep learning approaches were used to extract the spatiotemporal features and the image's quality indicators were evaluated against the mean absolute error. Our quality indicators encapsulate both anatomical and pathological elements to provide multivariate assessment scores for anatomical visibility, clarity, depth-gain and foreshortenedness.

Results The model performance accuracy yielded 94.4%, 96.8%, 96.2%, 97.4% for anatomical visibility, clarity, depth-gain and foreshortenedness, respectively. The mean model error of 0.375 ± 0.0052 with computational speed of 2.52 ms per frame (real-time performance) was achieved.

Conclusion The novel approach offers new insight to the objective assessment of transthoracic echocardiogram image quality and clinical quantification in A4C and PLAX views. It also lays stronger foundations for the operator's guidance system which can leverage the learning curve for the acquisition of optimum quality images during the transthoracic examination.

1. Introduction

A two-dimensional (2D) echocardiogram has become de facto of assessing cardiac functions because it presents rich anatomical details of the myocardium, and for its non-ionizing in-vivo advantages. Nevertheless, echocardiogram quality assessment is not void of technical and operational drawbacks. Firstly, echocardiogram images are produced through scattering centers and do not present crisp edges unlike the non-medical images. Secondly, the acquisition of high-quality echo images requires a significant experts' skill, and the standards of image quality are commonly exacerbated by user's subjective assessment and patients' anatomical profiles. For example, there is a strong indication why the quantification of systolic function is recommended for apical-four (A4C) and parasternal long axis (PLAX) views [1–3]. According to Nosir (1997) and Lang (2015), the spatial orientations of A4C and PLAX

views are congruent in orientation, thus offering complementary advantages on the heart's functional measurement and quantifications. However, echocardiogram quality assessment still exists in subjective domain, with significant impacts, as documented [4–6], ultimately pointing to the issues of clinical reliability, misdiagnosis, and poor response to patient care. These drawbacks remain significant and are inhibiting the adoption of echocardiograms as the reliable imaging modality despite its many advantages. Apparently, a good quality image provides a more accurate delineation of myocardial borders and yields accurate measurement [3], but the factor perceived as a 'good quality' indicator is based on individual assessment that varies throughout the clinical practices. This indicates the need for a coherent standard and benchmark for determining the constituents of a "good quality" image.

Currently, the method of echocardiographic image assessment entails manual inspection of echo images (sometimes large number of im-

* Corresponding author: Robert B. Labs, School of Computing and Engineering, University of West London, London, United Kingdom. (Email: robbie.labs@uwl.ac.uk)

<https://doi.org/10.1016/j.imed.2022.08.001>

Received 21 December 2021; Received in revised form 10 May 2022; Accepted 12 August 2022

2667-1026/© 2022 The Author(s). Published by Elsevier B.V. on behalf of Chinese Medical Association. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>)

ages) to determine its clinical and pathological relevance. This process is known to be time consuming, laboriously expensive, and precipitate variability of opinions on diagnostic outcomes. Consequently, an automated assessment is thus required for consistency, reliability, assessment, and optimization.

1.1. The objective of this study

The assessment of image quality in natural images (non-medical images) are quite straight forward with deep learning approaches, as they are modeled either with full-reference (FR-IQA) or reduce-reference (RR-IQA) metrics. On the other hand, 2D echocardiographic images are formed by an interference pattern of scattering centers with inherent poor lateral and axial resolutions. Basically, the echocardiogram's anatomical features do not present crisp edges and boundary either because the endocardium is trabeculated with papillary muscles and the external purkinje networks [7]. The complexity of the heart structure and the relationship between the epicardium, myocardium, and pericardium does exacerbate the acoustic impedance transition between its soft layers and precipitated the heterogeneity in echocardiograms. Because of this problem, it is grossly inadequate to model 2D echocardiograms under distortion-specific metrics [8]. The echocardiogram presents significantly subtle differences in its successive frames, that human eyes cannot detect [9]. This is a direct contrast to natural images, which usually feature well-known distortion types.

The objective of this study, therefore, focuses on 2D echocardiographic image quality indicators and the method of its assessment. Several research efforts have indicated the value of objective standards and proposed a number of assessment methods for evaluating echocardiographic image quality [10–12]. Unfortunately, the efforts do not meet translatory acceptance, either because they are limited in scope, or because they do not represent expert's clinical standard in objective benchmarking. Therefore, the implementation of an automated assessment protocol would be based on a fundamental definition of domain-specific objective standards and method of its assessment.

However, in the context of ultrasound, a coherent standard of objectivity relating to general echocardiography's digital image is essential but difficult to define outside the clinical practice [6]. But it is commonly admitted that the constituents of a good-quality echocardiogram image should be relative or congruent to clinical significance and anatomical delineation. This knowledge comes with many years of experts' professional experiences. Hence, an objective standard should include specific elements, that purposely encapsulate both the anatomical and pathological visibility, with the possibility for quality optimization and real-time assessment. Consequently, by defining the element of image quality attributes (indicators) in these two planes, we can effectively model an inclusive, objective standards benchmark that is relevant to generalized clinical protocols in echocardiography. We believe this would provide objective arbitration to improve reliability of cardiac measurement, quantification, and diagnostic accuracy in echocardiography.

1.2. Related work

All previously reported studies on objective assessment have used limited criteria to define objective quality and were deemed inappropriate of clinical assessment in clinical practice. The method of assessment using a weighted average score index has been considered unsuitable for real-time acquisition and optimization guidance. Hence, the practical deployability of such a system is limited to experimental demonstration instead of translatory advantage. A clinically relevant system therefore would provide insight on objective standard and method of assessing specific quality attributes.

We have earlier demonstrated the feasibility of such a system of assessment [7,13] and hereby provide details of wider implementations and how it can be clinically deployed in a unified workflow. Apart from validating an objective quality system with external clinical dataset, the

proof of concept for operators' guidance on automated echocardiography have been discussed in many studies [14–15], but implementing a useful pipeline for clinical advantages has achieved little impact. This is because a pipeline that offers weighted average scores is incapable of specificity in relation to the elements of image quality and operators' acquisition skills. All existing pipelines (without exception) that are capable of real-time feedback have only indicated the maximum quality ratings of echo images without suggesting 'how' or 'what' aspect of the image requires an improvement. This means that operators are left to utilize their acquisition experience, which is of a little benefit to less experienced operators.

One of the earliest works on objective assessment of cardiac image quality is Abdi et al. He demonstrated the feasibility of objective assessment using convolutional neural network models in five apical views using six criteria scoring methods [16–17]. Since there was no publicly available cardiac dataset to model, the author relied on expert's knowledge for its feature engineering, a high resource intensive process. Abdi's 85% of model accuracy was regarded as plausible outcome but was clinically deficient for transthoracic standard examination practice. This is because the defined quality indicators are limited and do not represent experts' global characteristics for cardiac diagnosis using 2D echocardiographic images.

Alternatively, Luong et al.'s research utilized twelve criteria to grade each of the nine apical standard views, while computing a continuous single variable score to represent objective quantity for respective apical views [18]. Luong's regression model achieved overall accuracy of 87% with regards to four expert ground truths and sufficiently demonstrated the impact of image quality in diagnostic utility. However, the assessment methods and scores do not represent cardiologists' conventional assessment in practice and hence, cannot be applied in clinical workflow.

The most recent study on objective quality assessment by Dong et al. [14]. However, the study was limited to fetal ultrasound in apical four-chamber plane (A4C) and did not include PLAX view nor similar score criteria that can be independently accessed for adults' echocardiography examination in clinical practice. Dong's argument for focus/zoom attributes emanated from fetal cardiology where specific tissue became the focus of an investigation. However, these image attributes, though important, should be described as elements of clarity. Therefore, a zoomed-in section of the myocardium should exhibit the attributes of clarity instead of being considered as an independent indicator.

1.3. Main contributions

In the light of the above related work, we admit the research efforts are plausible contributions, however, the specified criteria used to define quality assessment are limited in scope and are insufficient for TTE's clinical relevance. The existing assessment methods do not match expert's expectations as currently obtainable in clinical practice.

However, in this research, we examined all existing quality criteria and proposed additional criteria that could translate to experts' subjective assessment. Finally, we defined, for the first time, a novel, most comprehensive criteria, and objective attributes (quality indicators) by which cardiac images can be assessed and optimized. We summarize our main contributions as follows:

- Demonstrate the feasibility of a novel, coherent and clinically relevant objective standard for the assessment of 2D echocardiographic images that account for relevant anatomical profiles, linear and volumetric quantifications of myocardial functions.
- Fresh insight to real-time assessment method that provides access to specific elements of cardiac image quality for the purpose of image optimization, accurate quantification, and diagnosis.
- Annotation of an independent echocardiography patient dataset showing four attributes of image quality namely: anatomical visibility, chamber clarity, depth-gain, and fore-shortening attributes for A4C, PLAX apical standard views.

Table 1 View-specific scoring definition

A4C		PLAX	
Assessed element per attributes	Maximum manual scores awarded	Assessed element per attributes	Maximum manual scores awarded
Anatomical visibility		Anatomical visibility	
Correct axis, Apical segment	6	Left ventricle (LV) visible	5
Interventricular septum visible	2	Right ventricle (RV) visible	3
Interatrial septum visible	2	Full segment pericardium visible	2
Anatomical clarity		Anatomical clarity	
LV cavity clarity, clear edges	4	LV cavity clarity (distinguishable border)	4
Distinguishable valves	3	LV anteroseptal wall clarity	3
Distinguishable septum wall	3	LV inferolateral wall clarity	3
Signal depth-gain		Signal depth-gain	
Image sectorial gain	4	Sectorial gain	4
No excess gain	3	No excess gain	3
Minimum artefacts	3	Minimum artefacts	3
LV foreshorten		Cavity foreshorten	
LV apical segment present	4	No-apex diastole	5
Normal-shaped diastole	3	No-apex systole	5
Normal-shaped systole	3		

The quality of each view was evaluated according to several experts' elicited criteria; each criterion consisted of several attributes with independent scores but yielding a maximum score of 10 points for each criterion.

- Public release of experts' annotated patient dataset to allow future studies and external validation of the new approaches or methods available on request at IntSav-QLabs [31]
- Detailed implementation of multi-stream deep learning architecture pipeline to process and allow access to specific image attributes in A4C and PLAX view of echo cine loop.

2. Materials and methods

We provide a detailed account of the dataset description, and justification for the collective elements required for objective standard assessment of image quality on A4C and PLAX views, and how they can be optimized in real-time deployment. This is followed by the expert annotation process and details for the implementation of our deep convolutional neural network model.

2.1. Dataset source and ethical approval

At present, no echocardiogram dataset with the corresponding four separate annotations on A4C/PLAX image quality assessment is publicly available. We, therefore, aimed at preparing our own dataset (echocardiograms and corresponding ground-truth) for model developments. A large random sample of echocardiographic studies from different patients performed between 2010 and 2020 was extracted from Imperial College Healthcare NHS Trust's echocardiogram database. Ethical approval was obtained from the Health Regulatory Agency for the anonymized export of large quantities of imaging data. It was not necessary to approach patients individually for consent of data originally acquired for clinical purposes.

The images were acquired during examinations performed by experienced echocardiographers, according to the standard protocols for using ultrasound equipment from GE Healthcare and Philips Healthcare manufacturers. Automated anonymization was performed to remove the patient-identifiable information from DICOM-formatted videos.

A neural network model, previously developed in our research group [19,20], was then used to detect different echocardiographic views and separate the A4C and PLAX views. This resulted in a total of 33,784 frames from different patients: 15,476 and 18,308 frames for A4C and PLAX, respectively.

2.2. Definition and grouping of quality attributes indicators

View-specific image quality scoring indicators (attributes) and criteria was defined by consulting our clinical expert committee at the National Heart and Lung Institute. Four main quality attributes were

considered for each view, which are listed in Table 1, and enumerated as follows.

2.2.1. Anatomical visibility

Unlike photographic images, ultrasound images are formed by interference patterns of scattering centers that do not present clear edges, but inherently poor lateral and axial resolutions [7]. Hence, the magnitude of visibility on chamber cavities for both A4C and PLAX frames can be correctly visualized using the correct method of heart's apex slicing, to yield the acceptable clinical projection of images' anatomical structures. This could present a sharp or blurred edges [15] of amplitude structures. Equations (1 - 2) describe the rotation of a frame vector in 2D spatial distribution where x_1, y_1 represent on-axis projection, taking arbitrary center x_c, y_c , off-axis x_p, y_p can thus be mitigated from β known angle to improve anatomical visibility. In A4C, emphasis is placed on apical orientation, echogenicity of the left ventricle chamber, and mitral and atrium valves [2]. Although the LV apex is not visualized in PLAX, emphasis is placed on the anatomical echogenicity and clinical orientation of the right and left ventricles, the pericardium positions, and the aortic valves. These are clinically relevant features experts rely on for quantification, clinical assessment, and diagnosis.

$$x_1 = (x_p - x_c) \cos\beta - (y_p - y_c) \sin\beta + x_c \quad (1)$$

$$y_1 = (x_p - x_c) \sin\beta - (y_p - y_c) \cos\beta + y_c \quad (2)$$

2.2.2. Cavity clarity

Left ventricle clarity is a legacy attribute in objective assessment. Unlike non-medical images, apical chambers of any zoomed region can only present rough boundaries and contractive edges. Kurt, et al. [15] have demonstrated the impact of contrast echocardiography, however, with respect to quantification, cavity clarity is visualized by several distinguishable fast-moving pixel's formations during cardiac cycles. This attribute, therefore, addresses the degree of distinguishable pixel element representing the endocardial border cavities or clear distinction between the intraventricular septum, valves, any trabeculated pericardial fluids and endocardial walls. Cardiac frames with very high contrast or very low contrast represent the extreme end of the spectrum and pose significant challenges [4,21] to inexperienced operators. Equation (3) describes the root mean squared (RMS) contrast, C_i , which does not depend on angular frequency content or spatial distribution as best suited for 2D cardiac frames. This is given as the difference between the standard deviation of normalized pixel intensity $I_{i,j}$, and mean normalized intensity \hat{I} , of a given anatomical pathology; where

(i,j) represents the i -th and j -th element of 2D image size M, N ; An extreme contrast could generate artefacts and potentially obscured essential anatomical details. Unfortunately, echo images with low contrast do have significant anatomical details required for clinical measurement, suggesting the need to assess each image on the merits of clarity.

$$C_{i,j} = \left[\frac{1}{MN} \sum_{i=0}^{M-1} \sum_{j=0}^{N-1} (I_{i,j} - \hat{I})^2 \right]^{\frac{1}{2}} \quad (3)$$

2.2.3. Depth-gain

Depth-gain is peculiar to 2D echocardiography, and it represents a measure of intensity of discrete signal samples of a specific region of interest. The intensity of the image signals becomes susceptible to depth changes, sector width and patient's anatomical profile. Although the use of high frequency probes can yield better resolution at shallow tissue depth penetration [21], low frequency probes give the opposite effect. Consequently, signal gain at the image apex (near field) usually possesses strong intensity of high amplitude and could become excessively low at the far field region of the cardiac frame. In the same way, excessive gain can present as pulmonary fluid in some cases [13] and images with extremely low gain attributes but bear significant anatomical details or noticeable artefacts are considered in clinical practice. Equation (4) describes the intensity of the reflected beam, which is associated with depth gain; where $d2\phi$ represent the luminous flux of the infinitesimal area of source $d\Sigma$, divided by the product of $d\Sigma$, infinitesimal solid angle $d\Omega_\xi$ and $\theta\xi$ angle between the normal $\Omega\xi$ to the source $d\Sigma$. While the image luminance represents a photometric measure of a pixel's intensity per unit area of light for a given area of interest. Brightness therefore is the subjective impression of the object of luminance $I_{i,j}$ and is measured in candela per square meter cd/m^2 . The objective model therefore assesses and scores any potential introduction of artefacts from excessive gain, incorrect depiction of tissues or obscurity of relevant anatomical details that are relevant for measurements.

$$I_{i,j} = \frac{d^2\phi}{d\Sigma \cdot d\Omega_\xi \cos\theta_\xi} \quad (4)$$

2.2.4. Foreshortening

Apical foreshortening presents as a form of perspective deformation of the LV cavity, especially in the apex region. This deformation occurs as a result of poor image acquisition skills and could effectively alter the chamber's size and renders its volumes geometrically incongruent [22]. Apical foreshortening could occur during either the systolic or diastolic cycles, hence, both cycles are considered during the frame's real-time assessment. Smistad et al. have described the importance of real-time detection of apical foreshortening. For instance, foreshortening can result in inaccurate quantification of ejection fraction (EF) [14] or prevent the detection of crucial pathology, especially in the apical region. We refer to this undesirable perspective transformation $I_{(x,y,z)}$ which adds an additional layer z , to the image's 2D plane x,y is expressed in terms of the product of homogenous transformation properties given in Equation (5). In the PLAX view, however, where LV apex visibility is not required, visible apex of the LV could be taken as 'false-apex' [2], therefore counts as LV foreshortening. From a clinical standpoint, eliminating foreshortenedness is paramount to optimal quantification, anatomical assessment and diagnosis of many ailments including cardiomyopathy.

$$I_{x,y,z} = \begin{bmatrix} 1 & 0 & 0 & 0 \\ 0 & 1 & 0 & 0 \\ 0 & 0 & 1 & 0 \\ 0 & 0 & -\frac{1}{d} & 1 \end{bmatrix} \begin{bmatrix} x \\ y \\ z \\ 1 \end{bmatrix} = \begin{bmatrix} x \\ y \\ z \\ -\frac{1}{d} \end{bmatrix} \Rightarrow \left(-d \frac{x}{z}, d \frac{y}{z} \right) \quad (5)$$

2.3. Expert annotations

To establish the ground-truth scoring for neural network developments and testing, each echo video underwent annotation process by

an experienced expert, who provided an independent score value for each quality attribute defined in Table 1. The score ranges from 0 to 9 under each attribute, to allow for specificity and fair assessment of the A4C/PLAX apical standard. Therefore, the multi-stream architecture was trained on all four attributes simultaneously to provide normalized objective scores in the final output. Expert annotations for the echo videos were used as the quality score for all constituent frames of that video for the model developments.

2.4. Dataset preparations

The study population consisted of a random sample of 11,262 echographic studies from patients with age ranges from 17 to 85 years, who were recruited from patients who had undergone echocardiography. Three frames were randomly drawn from the video and split into training (27,028 frames), and testing (6,756 frames) sub-datasets in 80:20 ratios. Figure 1 summarizes the frame distributions for A4C and PLAX with a categorical characteristic using experts' maximum score range values of 4.5, 6.5, 9.9 designated as poor, average, and good quality (Figure 2) respectively. Any image with 0 score was rejected and considered unsuitable for model development.

2.5. Neural network architecture

The architecture used in this study, referred to as 'QA-NET' is based on a multi-stream, multi-output regression model, featuring four sub-node model architectures fused together to simultaneously train and make predictions in a multi-labelled fashion. The model accepts input frame of variable length (spatial size 224×224 pixels) indicated by weight matrix w_i^l and convolved with the convolution layer of each parallel sub-model F^{i-1} as the input feature-map to achieve a 2D output feature map, F_i^l of i^{th} kernel of the specific convolution layer l , given in Equation (6). This flattened vector is fed into a layer of time sequence module (LSTM) for temporal extraction. Each convolution layer features an activation function of the type - Rectifier Linear Units (ReLU) [23], as shown in Equation (7) where neurons' activation values of x ranges from 0 to maximum value. Each node was dedicated to extracting a specific anatomical feature relating to criteria defined for A4C and PLAX standard views in Table 1. The components of the sub-node architecture are structurally optimized for each specific quality attribute and adapted based on best performing architecture against each quality attribute. The model features both spatial and temporal modules, illustrated in Figure 3, and detailed as follows:

- Each of the spatial module consists of four convolutional layers (except the clarity module with three convolutional layers), with kernel configuration of 32, 32, 32, 64 and of 3×3 size, respectively. Each convolutional layer features batch normalization [24], except the third convolutional layer which missed out on max pooling [25], and dropout of 0.5 [26]. The output is flattened, and the sequence is fed into the temporal module.
- Temporal module consists of an LSTM layer, used to extract temporal features. It accepts vector data from each adjacent convolutional module to compute mean score on frames' sequence data. The features are based on fast changing pixel intensity between consecutive frames, which resulted in noises where increase in vanishing gradients on training data became apparent. Therefore, output layers were conFig.d differently to feature two stages of dense layer, batch normalization and dropouts of 0.5. This was noted to offer resilience against noisy labels and reduce variance in image/frame data.

The choice of architecture was based on the performance data, memory requirement, and fastest inference speed data, all of which are significant for real-time feedback implementation. We also investigated of the well-known, state-of-the-art network architectures as found in relevant




Figure 1. Summary of data distribution for A4C and PLAX cardiac frames, indicating three categories of quality levels based on experts scores values: Frames with max scores of 4.5, 6.5 and 9.9, are classified as Poor quality, Average quality, and Good quality, respectively.




Figure 2. Samples of A4C with (a) cut-off apex, (b) slightly foreshortened, and (c) good-quality image with clear visibility of interventricular septum, the left ventricle (LV) correctly projected (on-axis), depth-gain, and minimal foreshortening would likely gain high prediction values in quality assessment.




Figure 3. The multi-stream neural network architecture referred to as 'Q-NET'. Each stream is dedicated to specific prediction and assessment of images quality on visibility, clarity, depth-gain and foreshortenedness as identified by QV, QC, QD, and QF, respectively.

resources: DenseNet121 [27], ResNet [28], VggNet [29] and compare the performance along with each 2d+t hybrid versions.

The model was trained using a 5-fold cross validation technique to ensure adequate learning on the dataset and performance was recorded for each model. The hyper parameters learning rate was set at 0.002 with a high momentum of 0.95 and decay rate of 0.1 every 24 steps, and they were reproducibly initialized to minimize possible deviations in score performance. Data augmentation was applied to allow optimum learning sequences for the models; a maximum translation of [-0.05, +0.05] pixels and maximum rotation of 5 degrees were applied randomly for horizontal, vertical, and rotational angles, respectively. To prevent overfitting in the training phase, we applied batch normalization and dropout. A multi-label optimization approach was adopted [30], and the model was trained simultaneously using four quality attributes with mean absolute error as the cost function. Training was initialized with 32 batch size and completed as learning curves converged

around 40 epochs.

$$F_{i,j,k}^l = \sum_{i=0}^n \sum_{j=0}^m w_{i,mn}^l F_{(j+m)(k+n)}^{l-1} \quad (6)$$

$$f(x)_{relu} = \max(0, x) \quad (7)$$

2.6. Evaluation metrics

Since the model uses multiplex variables for each score attributes, the output score was normalized to [0, 1] via sigmoid activation function $f(x)$, Equation (8) and prediction error were evaluated against the MAE, taking the average of the absolute difference between cardiologist's ground truth (Q_{GT}) scores and model's predicted scores (Q_p). Therefore, model's minimal error with values close to 0 would indicate a best fit scenario, while a larger error with values close to 1 would indicate a poor fit regression model. Lastly, the average model's performance in percentage was computed in Equation (9).

$$f(x)_{sigmoid} = 1/(1 + e^{-x}) \quad (8)$$

$$Model_{acc} = 1 - \left(\frac{\sum_{i=0}^n |Q_{GTi} - Q_{pi}|}{n} \right) * 100 \quad (9)$$

3. Results

Cardiac echo frames are laced with significant complexities, among which are patients' anatomical and pathological differences ; these complexities are reflected in each fast-moving echo frame, therefore, the model's inference speed is overly critical to real-time assessment and operators' feedback guidance for possible optimization. Hence, implementing a customized model that can successfully generalize with high confidence and high-speed inference would make a significant achievement in automated assessment.

Table 2 Model performance on quality attribute/indicator and the error distribution, given by the mean and deviation ($\mu \pm \sigma$) notation

Quality model / indicator	Accuracy (%)	Q1	Q2	Q3	Inference time (ms)
		$\mu \pm \sigma$			
Visibility	94.37	0.1536 ± 0.1036	0.1681 ± 0.1100	0.1581 ± 0.1038	9.476
Clarity	96.84	0.1635 ± 0.1034	0.1759 ± 0.1042	0.1837 ± 0.1106	7.753
Depth-gain	96.27	0.1357 ± 0.0982	0.1326 ± 0.0956	0.1306 ± 0.0956	8.384
Foreshorten	97.50	0.1909 ± 0.1262	0.1901 ± 0.1285	0.2044 ± 0.1242	9.356

Q1, Q2, and Q3 represent the three levels of image quality per model. Inference time is the average time it takes the model to identify and predict scores for each image's quality levels.




Figure 4. Box plot shows the error distribution on the multi-stream model architecture featuring the collective quality attributes indicators. The y-axis shows the difference between experts' ground truth and model predictions. The x-axis shows each quality attributes (indicator) per model.

The proposed multi-stream model was evaluated on an external dataset to lower systemic bias and achieved a mean accuracy of 96.20% and 2.52 ms inference speed which reinforces the viability for real-time feedback deployment per quality per frame. The results enumerated in Table 2 outlined the accuracies and error distribution achieved on each dedicated model for visibility, clarity, depth-gain and foreshorten indicators respectively. The aggregated error distribution per model is de-

picted in Figure 4, while the model predictive accuracy (Table 3) for visibility, clarity, depth-gain, and foreshortening group attributes (indicators) are 94.4%, 96.8%, 96.2% and 97.4% respectively. The samples of predicted cardiac frames shown in Figure 5 that clearly indicates the predictive objective scores for visibility, clarity, depth-gain, foreshortenedness and its weighted average score (AS). These are automatically generated by the pipeline and superimposed on the cardiac frames in real-time.

4. Discussion

The results of each model's performance (except QA-NET+LSTM) are shown in Table 3, and vary substantially by group attribute/indicator even though each model retains its original values of hyperparameters. This indicatively prove that one model cannot fit it all. Each of the evaluated state-of-the-art model is incapable of delivering consistent performance in terms of accuracy, and real-time inference speed on the selected group quality attributes/indicators.

The achieved inference speeds for DenseNet121, ResNet50, and Vg-gNet16, are 24.70ms, 19.53ms, and 30.76ms, respectively. This implies that a maximum frame per second (FPS) of 40, 51 and 32 can be achieved with the respective state-of-the-art models. This speed is insufficient for real-time quality assessment. Nevertheless, our model (QA-Net) achieved a reduction of 90% in inference speed for the combined




Figure 5. Sample of predicted images with respective objective scores. Visibility (VS), clarity (LC), depth-gain (DG), foreshortening (FS) and overall quality score (AS) are used to assess cardiac image quality during clinical acquisition process.

Table 3 Comparison of the model’s performance, including the selected state-of-the-art model (DenseNet, ResNet and VggNet) on echocardiogram objective quality score

Model data	Regression accuracy (%)					Inference time (ms)
	Visibility	Clarity	Depth-gain	Foreshorten	Accuracy	
DENSENET+LSTM	92.20	88.20	95.62	90.44	91.62	24.704
RESNET+LSTM	89.45	92.25	87.40	92.20	90.32	19.526
VGGNET+LSTM	92.30	97.20	98.40	89.20	94.28	30.760
QA-NET+LSTM	94.40	96.80	96.20	97.40	96.20	2.520

Each of the models was evaluated on the combined attributes (quality indicator) of visibility, clarity, depth-gain, and foreshortening simultaneously. The best performing model (QA-Net), in terms of accuracy and inference is highlighted.




Figure 6. Illustrates the convolutional layers learned features (feature map) from a specific image (original input image) and its respective layers of distinguishable detection from each model.

model performance, making it the best candidate for transthoracic image quality assessment solution. Note that our multi-stream pipeline actively combined four attributes group indicators simultaneously during the training and prediction phases. The deep neural networks were fully optimized to yield the best inference speed and performance accuracy on each of the specific quality indicator. Therefore, we concluded that each quality attribute/indicator associated with distinguishable pathological complexities and fast-changing echocardiogram would require a fully customized and optimized model for objective quality assessment task.

Furthermore, a summary of visualization of the learned features is illustrated in Figure 6. To obtain this, each quality indicator group was modeled individually to show the discriminative ability of the respective network of our model. Here, the model’s feature maps give the idea on which part of image’s element is being focused upon at each convolutional layer in the network. Although several global characteristics have been used in the quality criteria, this study does not claim exhaustiveness in the group criteria indicators. We are aware that different laboratories are at liberty to adopt what is considered the best practice in their region of practice, especially when such requirements are mandatory by healthcare legislation.

5. Conclusion

In this study, we considered four distinctive group of quality attributes or indicators where echocardiographic image quality was eval-

uated and a novel method of accessing such attributes under A4C and PLAX echocardiographic views. These two views bear clinical significance and are recommended for chamber quantification and linear measurement in clinical workflow [1–3]. Since this appear to be the first time where 2D echocardiogram quality indicators are thus defined comprehensively, the results of our work can only be compared to the recent, and existing work on echocardiographic image quality assessment listed in Table 4. Technical comparison, in terms of the use case functionality, clinical feasibility, and method of assessment could provide significant evidence on how this important clinical problem in echocardiographic image quality has been addressed. Furthermore, each of the existing work (including ours) has employed different independent dataset, study population and sample size. Even though our works achieved better model accuracy compared to all the existing similar works, it is reasonably unfair to compare the model accuracy or inference speed event, except for clinical deployability and use case functionality.

We have presented the clinical significance and feasibility of developing an automated quality assessment in 2D echocardiographic images. Therefore, a quantitative method defined for image quality standard can provide useful feedback for an operator guidance system and a valuable tool for research in clinical practice, first to function as an arbiter reference to clinicians and, secondly, to accelerate the learning curve for those in training. Also, it can provide specific information on the adequacy of the images obtained in retrospect, which could be universally relevant for a lifesaving procedure at the point of care or during clinical emergencies.

Table 4 Summary of model performance on quality assessment pipeline in related studies

Studies	Ultrasound source	Study population (n)	Ground truth annotations	Quality criteria	Of image quality attributes and assessment methods	Of standard views considered and type	Input size	Sample size (n)	Model accuracy achieved (%)
Abdi et al., (2017) [16]	Philips and GE	N/A	2 expert annotations	13	1 (WAS)	5 (A2C, A3C, A4C, PSAXA, PSAXPM)	200 × 200 × 3	6,916	85.00
Luong et al., (2020) [18]	Philips iE33 platform, Philips S51 frequency 5–1 MHz	3,157	1 level 3 echo cardiographer	12	1 (WAS)	9 (PLAX, A2C, A3C, A4C, PSAX-A, PSAX-M, PSAX-PM, SC4 & IV(C))	Not specified	14,086	87.00
Dong et al., (2020) [14]	Shenzhen Maternal and Child Healthcare Hospital	N/A	1 radiological Expert	6	2 (Zoom, Gain)	1 (A4C only)	224 × 224 × 3	2,032	93.52
Current Study (2022)	GE Healthcare (Vivid.i) and Philips Healthcare (iE33 xMATRIX)	11,262	4 annotations each by 2 AV experts	23 (Table 1)	4 (Visibility, Clarity, Depth-Gain & Foreshortening)	2 (A4C & PLAX)	224 × 224 × 3	33,784	96.20

Finally, we used the annotation provided by two experts; a cardiologist who provided reference and supervision and an accredited annotator. An intra-observer variability can be examined by obtaining additional annotations from human experts and compared them to the error in the predicted scores.

Conflicts of interest statement

The authors declare that there are no conflicts of interest.

Funding

This research did not receive any specific grant from funding agencies in the public, commercial, or not-for-profit sectors.

Author contributions

Robert Labs: original draft preparation, data curation, writing; **Apostolos Vrettos:** ground truth annotations; **Jonathan Loo:** validation, reviewing; **Massoud Zolgharni:** reviewing, formatting, editing.

References

- Nosir YFM, Vletter WB, Boersma E, et al. The apical long-axis rather than the two-chamber view should be used in combination with the four-chamber view for accurate assessment of left ventricular volumes and function. *Eur Heart J* 1997;18:1175–85. doi:10.1093/oxfordjournals.eurheartj.a015414.
- Mitchell C, Rahko PS, Blauwet LA, et al. Guidelines for performing a comprehensive transthoracic echocardiographic examination in adults: recommendations from the american society of echocardiography. *J Am Soc Echocardiogr* 2019;32:1–64. doi:10.1016/j.echo.2018.06.004.
- Lang RM, Badano LP, Mor-Avi V, et al. Recommendations for cardiac chamber quantification by echocardiography in adults: an update from the american society of echocardiography and the european association of cardiovascular imaging. *Eur Heart J Cardiovasc Imaging* 2015;16:233–71. doi:10.1093/ehjci/jev014.
- Nagata Y, Kado Y, Onoue T, et al. Impact of image quality on reliability of the measurements of left ventricular systolic function and global longitudinal strain in 2D echocardiography. *Echo Res Pract* 2018;5:27–39. doi:10.1530/ERP-17-0047.
- Liao Z, Girgis H, Abdi A, et al. On Modelling Label Uncertainty in Deep Neural Networks: Automatic Estimation of Intra-observer Variability in 2D Echocardiography Quality Assessment; 2019. ArXiv1911.00674 Cs Eess Stat.
- Sprawls P. Optimizing Medical Image Contrast, Detail and Noise in the Digital Era; 2014.
- Labs RB, Vrettos A, Azarmehr N, et al. ICRMIRO 2020 Int Conf Radiol Med Imaging Radiat Oncol Paris; 2020.
- Sun S, Yu T, Xu J, et al. GraphIQA: Learning Distortion Graph Representations for Blind Image Quality Assessment; 2022. ArXiv210307666 Cs.
- Zhu L, Xu Z, Fang T. Analysis of cardiac ultrasound images of critically ill patients using deep learning. *J Healthc Eng* 2021;2021:1–8. doi:10.1155/2021/6050433.
- Rosen AFG, Roalf DR, Ruparel K, et al. Quantitative assessment of structural image quality. *NeuroImage* 2018;169:407–18. doi:10.1016/j.neuroimage.2017.12.059.
- Gaudet J, Waechter J, McLaughlin K, et al. Focused critical care echocardiography: development and evaluation of an image acquisition assessment tool. *Crit Care Med* 2016;44:e329–35. doi:10.1097/CCM.0000000000001620.
- Huang KC, Huang CS, Su MY, et al. Artificial intelligence aids cardiac image quality assessment for improving precision in strain measurements. *JACC Cardiovasc Imaging* 2021;14:335–45. doi:10.1016/j.jcmg.2020.08.034.
- Labs RB, Zolgharni M, Loo JP. Proceedings of 25th Annual Conference on Medical Image Understanding and Analysis. Cham: Springer International Publishing; 2021. doi:101007/978-3-030-80432-9_36.
- Dong J, Liu S, Liao Y, et al. A generic quality control framework for fetal ultrasound cardiac four-chamber planes. *IEEE J Biomed Health Inform* 2020;24:931–42. doi.org/10.1109/JBHI.2019.2948316.
- Smistad E, Ostvik A, Salte IM, et al. Real-time automatic ejection fraction and foreshortening detection using deep learning. *IEEE Trans Ultrason Ferroelectr Freq Control* 2020;1–1. doi:10.1109/TUFFC.2020.2981037.
- Abdi AH, Luong C, Tsang T, et al. Proceedings of 20th International Conference on Medical Image Computing and Computer-Assisted Intervention. Cham: Springer International Publishing; 2017. doi:101007/978-3-319-66179-7_35.
- Abdi AH, Luong C, Tsang T, et al. Automatic quality assessment of echocardiograms using convolutional neural networks: feasibility on the apical four-chamber view. *IEEE Trans Med Imaging* 2017;36:1221–30. doi:10.1109/TMI.2017.2690836.
- Luong C, Liao Z, Abdi A, et al. Automated estimation of echocardiogram image quality in hospitalized patients. *Int J Cardiovasc Imaging* 2021;37:229–39. doi:10.1007/s10554-020-01981-8.
- Azarmehr N, Ye X, Howard JP, et al. Neural architecture search of echocardiography view classifiers. *J Med Imaging* 2021;8. doi:10.1117/1.JMI.8.3.034002.
- Kurt M, Shaikh KA, Peterson L, et al. Impact of contrast echocardiography on evaluation of ventricular function and clinical management in a large prospective cohort. *J Am Coll Cardiol* 2009;53:802–10. doi:10.1016/j.jacc.2009.01.005.

- [21] Sassaroli E, Crake C, Scorza A, et al. Image quality evaluation of ultrasound imaging systems: advanced B-modes. *J Appl Clin Med Phys* 2019;20:115–24. doi:[10.1002/acm2.12544](https://doi.org/10.1002/acm2.12544).
- [22] Ünlü S, Duchenne J, Mirea O, et al. Impact of apical foreshortening on deformation measurements: a report from the EACVI-ASE strain standardization task force. *Eur Heart J Cardiovasc Imaging* 2019;jez189. doi:[10.1093/ehjci/jez189](https://doi.org/10.1093/ehjci/jez189).
- [23] Nair V, Hinton GE. *Proceedings of International Conference on International Conference on Machine Learning*. Omnipress; 2010.
- [24] Ioffe S, Szegedy C. *Batch Normalization: Accelerating Deep Network Training by Reducing Internal Covariate Shift*; 2015.
- [25] Wu H, Gu X. *Proceedings of International Conference on International Conference on Machine Learning*. Cham: Springer International Publishing; 2015. doi:[10.1007/978-3-319-26532-2_6](https://doi.org/10.1007/978-3-319-26532-2_6).
- [26] Srivastava N, Hinton G, Krizhevsky A, et al. *Proceedings of 29th IEEE Conference on Computer Vision and Pattern Recognition*, Las Vegas, United States; 2016. doi:[10.1109/CVPR.2016.90](https://doi.org/10.1109/CVPR.2016.90).
- [27] Huang G, Liu Z, van der Maaten L, et al. *Densely Connected Convolutional Networks*. IEEE Computer Society; 2016. doi:[10.1109/CVPR2017243](https://doi.org/10.1109/CVPR2017243).
- [28] He K, Zhang X, Ren S, et al. *Deep Residual Learning for Image Recognition*. *Proceedings of 2016 IEEE Conference on Computer Vision and Pattern Recognition (CVPR)*; 2016. doi:[10.1109/cvpr201690](https://doi.org/10.1109/cvpr201690).
- [29] Simonyan K, Zisserman A. *Very Deep Convolutional Networks for Large-Scale Image Recognition*; 2015. doi:[10.48550/arXiv.1409.1556](https://doi.org/10.48550/arXiv.1409.1556).
- [30] Díez J, Luaces O, del Coz JJ, et al. *Optimizing different loss functions in multilabel classifications*. *Prog Artif Intell* 2015;3:107–18. doi:[10.1007/s13748-014-0060-7](https://doi.org/10.1007/s13748-014-0060-7).
- [31] IntSavQ-Labs. Digital repository for IntSav projects. Available from <https://github.com/intsav> (Accessed on 27 Feb 2022).